Manufacturer Disclosure Statement for Medical Device Security -- MDS2 TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMAX Carestream Health. Inc. TX65 Laser Imager AN8943 1-Feb-2024 Question ID Question See note DOC-1 Manufacturer Name Carestream Health, Inc. DOC-2 Device Description TRIMAX Laser Imaging System TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and DOC-3 Device Model TRIMAX TX65 Laser Imager Document ID AN8943 DOC-4 Technical Support, Company Name: Carestream Health, Inc. -DOC-5 Manufacturer Contact Information Phone #: 1-800-328-2910 1. Prints medical film for diagnosis to determine patient treatment 2. Prints medical film to share with patients and referring physicians for education and assist with patient treatment 3. Prints medical film for archiving to compare current and previous medical images to monitor the progression of treatment 4. Prints affordable and portable medical film for patients that DOC-6 Intended use of device in network-connected environment: will resist the test of time and technology obsolescence 2/1/2024 DOC-7 Document Release Date Coordinated Vulnerability Disclosure: Does the manufacturer have a vulnerability disclosure program for this device? DOC-8 Yes ISAO: Is the manufacturer part of an Information Sharing and DOC-9 Analysis Organization? Yes Diagram: Is a network or data flow diagram available that indicates connections to other system components or expected external DOC-10 resources? Yes SaMD: Is the device Software as a Medical Device (i.e. software-DOC-11 only, no hardware)? No DOC-11.1 Does the SaMD contain an operating system? N/A Does the SaMD rely on an owner/operator provided operating system? DOC-11.2 N/A DOC-11.3 Is the SaMD hosted by the manufacturer? N/A DOC-11.4 Is the SaMD hosted by the customer? N/A Yes, No, N/A, or See Note Note # MANAGEMENT OF PERSONALLY IDENTIFIABLE INFORMATION Can this device display, transmit, store, or modify personally identifiable information (e.g. electronic Protected Health Store only information as annotations as a part of a MPII-1 Information (ePHI))? Yes **DICOM** image Data files that are obtained from the connected modality may include ePHI data burned in to the image. Image files are deleted from memory after MPII-2 Does the device maintain personally identifiable information? the print job is printed. No Does the device maintain personally identifiable information temporarily in volatile memory (i.e., until cleared by power-off or MPII-2.1 reset)? N/A Does the device store personally identifiable information MPII-2.2 persistently on internal media? N/A Is personally identifiable information preserved in the device's non-MPII-2.3 volatile memory until explicitly erased? N/A Does the device store personally identifiable information in a database? MPII-2.4 N/A Does the device allow configuration to automatically delete local personally identifiable information after it is stored to a long term MPII-2 5 solution? N/A Does the device import/export personally identifiable information with other systems (e.g., a wearable monitoring device might export MPII-2.6 personally identifiable information to a server)? N/A Does the device maintain personally identifiable information when MPII-2.7 powered off, or during power service interruptions? N/A Does the device allow the internal media to be removed by a service technician (e.g., for separate destruction or customer MPII-2.8 retention)? Yes Does the device allow personally identifiable information records be stored in a separate location from the device's operating system (i.e. secondary internal drive, alternate drive partition, or remote MPII-2 9 storage location)? N/A

MPII-3.2 Does the device have mechanisms used for the transmitting, Does the device of personally identifiable information? No Does the device display personally identifiable information (e.g., MPII-3.1 video display, etc.)? No Does the device generate hardcopy reports or images containing MPII-3.2 personally identifiable information? No

PII may be displayed as annotions on a hardcopy film print output by the imager.

Manufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers AN8943 A Uncontrolled unless otherwise indicated

	Does the device retrieve personally identifiable information from or	
	record personally identifiable information to removable media (e.g.,	
	removable-HDD, USB memory, DVD-R/RW,CD-R/RW, tape, CF/SD	
MPII-3.3	card, memory stick, etc.)?	No
	Does the device transmit/receive or import/export personally	
	identifiable information via dedicated cable connection (e.g., RS-	
MPII-3.4	232, RS-423, USB, FireWire, etc.)?	No
	Does the device transmit/receive personally identifiable information	
MPII-3.5	via a wired network connection (e.g., RJ45, fiber optic, etc.)?	Yes
	Does the device transmit/receive personally identifiable information	
	via a wireless network connection (e.g., WiFi, Bluetooth, NFC,	
MPII-3.6	infrared, cellular, etc.)?	No
	Does the device transmit/receive personally identifiable information	
MPII-3.7	over an external network (e.g., Internet)?	No
	Does the device import personally identifiable information via	
MPII-3.8	scanning a document?	No
	Does the device transmit/receive personally identifiable information	
MPII-3.9	via a proprietary protocol?	No
	Does the device use any other mechanism to transmit, import or	
MPII-3.10	export personally identifiable information?	No

Management of Private Data notes:

AUTOMATIC LOGOFF (ALOF)

The device's ability to prevent access and misuse by unauthorize	2d
users if device is left idle for a period of time.	

	Can the device be configured to force reauthorization of logged-in	
	user(s) after a predetermined length of inactivity (e.g., auto-logoff,	
ALOF-1	session lock, password protected screen saver)?	N/A
	Is the length of inactivity time before auto-logoff/screen lock user or	
ALOF-2	administrator configurable?	N/A

AUDIT CONTROLS (AUDT)

The ability to reliably audit activity on the device.

	Can the medical device create additional audit logs or reports	
AUDT-1	beyond standard operating system logs?	Yes
AUDT-1.1	Does the audit log record a USER ID?	Yes
	Does other personally identifiable information exist in the audit	
AUDT-1.2	trail?	No
	Are events recorded in an audit log? If yes, indicate which of the	
AUDT-2	following events are recorded in the audit log:	Yes
AUDT-2.1	Successful login/logout attempts?	Yes
AUDT-2.2	Unsuccessful login/logout attempts?	Yes
AUDT-2.3	Modification of user privileges?	No
AUDT-2.4	Creation/modification/deletion of users?	No
AUDT-2.5	Presentation of clinical or PII data (e.g. display, print)?	Yes
AUDT-2.6	Creation/modification/deletion of data?	No
	Import/export of data from removable media (e.g. USB drive,	
AUDT-2.7	external hard drive, DVD)?	N/A
	Receipt/transmission of data or commands over a network or point-	
AUDT-2.8	to-point connection?	Yes
AUDT-2.8.1	Remote or on-site support?	Yes
AUDT-2.8.2	Application Programming Interface (API) and similar activity?	N/A
AUDT-2.9	Emergency access?	N/A
AUDT-2.10	Other events (e.g., software updates)?	Yes
AUDT-2.11	Is the audit capability documented in more detail?	No
	Can the owner/operator define or select which events are recorded	
AUDT-3	in the audit log?	No
	Is a list of data attributes that are captured in the audit log for an	
AUDT-4	event available?	No
AUDT-4.1	Does the audit log record date/time?	Yes
	Can date and time be synchronized by Network Time Protocol (NTP)	
AUDT-4.1.1	or equivalent time source?	Yes
AUDT-5	Can audit log content be exported?	Yes
AUDT-5.1	Via physical media?	No
AUDT-5.2	Via IHE Audit Trail and Node Authentication (ATNA) profile to SIEM?	No
	Via Other communications (e.g., external service device, mobile	
AUDT-5.3	applications)?	Yes
AUDT-5.4	Are audit logs encrypted in transit or on storage media?	No
AUDT-6	Can audit logs be monitored/reviewed by owner/operator?	No
AUDT-7	Are audit logs protected from modification?	No
AUDT-7.1	Are audit logs protected from access?	Yes
AUDT-8	Can audit logs be analyzed by the device?	No

_
Image data with PII as annotations can be received in DICOM image format over Ethernet cable.
-
-

The c	levice does not allow a user to logon to th
syste	m.

Only service users may log in
Only service users may log in
_
—
_
_
_
_
Log files access by service
-
Audit logs are accessibly by Carestream authorized
Service personnel
Service personner

AUTHORIZATION (AUTH)

Manufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers AN8943 A Uncontrolled unless otherwise indicated

1-Feb-2024

Carestream Health, Inc.	TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMA> TX65 Laser Imager	(AN8943	1-Feb-2024
	The ability of the device to determine the authorization of users.		
AUTH-1	Does the device prevent access to unauthorized users through user login requirements or other mechanism?	N/A	New users cannot be created
AUTH-1.1	Can the device be configured to use federated credentials management of users for authorization (e.g., LDAP, OAuth)?	N/A	New users cannot be created
AUTH-1.2	Can the customer push group policies to the device (e.g., Active Directory)?	N/A	New users cannot be created
AUTH-1.3	Are any special groups, organizational units, or group policies	N/A	New users expect he created
A01H-1.5	required? Can users be assigned different privilege levels based on 'role' (e.g.,	N/A	New users cannot be created
AUTH-2	user, administrator, and/or service, etc.)?	N/A	New users cannot be created
AUTH-3	Can the device owner/operator grant themselves unrestricted administrative privileges (e.g., access operating system or application via local root or administrator account)?	N/A	New users cannot be created
AUTH-4	Does the device authorize or control all API access requests?	Νο	Service access is authenticated
	Does the device run in a restricted access mode, or 'kiosk mode', by		
AUTH-5	default?	Yes	
	CYBER SECURITY PRODUCT UPGRADES (CSUP) The ability of on-site service staff, remote service staff, or authorized customer staff to install/upgrade device's security patches.		
	Does the device contain any software or firmware which may require security updates during its operational life, either from the device manufacturer or from a third-party manufacturer of the software/firmware? If no, answer "N/A" to questions in this		
CSUP-1	section.	Yes	<u> </u>
CSUP-2	Does the device contain an Operating System? If yes, complete 2.1-2.4.	Yes	
			_
CSUP-2.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	No	
	Does the device require vendor or vendor-authorized service to		
CSUP-2.2	install patches or software updates? Does the device have the capability to receive remote installation of	Yes	—
CSUP-2.3	patches or software updates?	Yes	_
CSUP-2.4 CSUP-3	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer? Does the device contain Drivers and Firmware? If yes, complete 3.1- 3.4.	No Yes	_
	Does the device documentation provide instructions for		
CSUP-3.1	owner/operator installation of patches or software updates?	No	_
CSUP-3.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	Yes	
	Does the device have the capability to receive remote installation of		
CSUP-3.3	patches or software updates?	Yes	-
CSUP-3.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer? Does the device contain Anti-Malware Software? If yes, complete	Νο	_
CSUP-4	4.1-4.4.	No	—
CSUP-4.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	_
CSUP-4.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	
CSUP-4.3	Does the device have the capability to receive remote installation of patches or software updates?	N/A	
0.01-4.5	patches of software updates:	1/6	—
CSUP-4.4	Does the medical device manufacturer allow security updates from any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A	_
CSUP-5	Does the device contain Non-Operating System commercial off-the- shelf components? If yes, complete 5.1-5.4.	Yes	-
	Does the device documentation provide instructions for		
CSUP-5.1	owner/operator installation of patches or software updates? Does the device require vendor or vendor-authorized service to	No	—
CSUP-5.2	install patches or software updates?	Yes	_
CSUP-5.3	Does the device have the capability to receive remote installation of patches or software updates?	Yes	
CSUP-5.4	any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	No	

No

without approval from the manufacturer?

CSUP-5.4

Carestream Health, Inc.	TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMAX TX65 Laser Imager	AN8943	1-Feb-2024
	Does the device contain other software components (e.g., asset		
CSUP-6	management software, license management)? If yes, please provide details or refernce in notes and complete 6.1-6.4.	No	_
CSUP-6.1	Does the device documentation provide instructions for owner/operator installation of patches or software updates?	N/A	_
CSUP-6.2	Does the device require vendor or vendor-authorized service to install patches or software updates?	N/A	_
CSUP-6.3	Does the device have the capability to receive remote installation of patches or software updates?	N/A	_
	Does the medical device manufacturer allow security updates from		
CSUP-6.4	any third-party manufacturers (e.g., Microsoft) to be installed without approval from the manufacturer?	N/A	_
CSUP-7	Does the manufacturer notify the customer when updates are approved for installation?	Yes	If updates are mandataory
CSUP-8	Does the device perform automatic installation of software updates?	Yes	For RMS Smart Link connected devices
CSUP-9	Does the manufacturer have an approved list of third-party software that can be installed on the device?	No	
	Can the owner/operator install manufacturer-approved third-party		-
CSUP-10	software on the device themselves?	No	-
CSUP-10.1	Does the system have mechanism in place to prevent installation of unapproved software?	No	Software could only be installed when the user has heightened priviledges as a Service representative
CSUP-11	Does the manufacturer have a process in place to assess device vulnerabilities and updates?	Yes	
CSUP-11.1	Does the manufacturer provide customers with review and approval status of updates?		—
CSUP-11.1 CSUP-11.2	Is there an update review cycle for the device?	Yes	-
	HEALTH DATA DE-IDENTIFICATION (DIDT) The ability of the device to directly remove information that allows identification of a person.		
DIDT-1	Does the device provide an integral capability to de-identify personally identifiable information?	No	
DIDT-1.1	Does the device support de-identification profiles that comply with the DICOM standard for de-identification?	No	_
			_
	DATA BACKUP AND DISASTER RECOVERY (DTBK) The ability to recover after damage or destruction of device data,		
	hardware, software, or site configuration information.		
DTBK-1	Does the device maintain long term primary storage of personally identifiable information / patient information (e.g. PACS)?	Νο	
	Does the device have a "factory reset" function to restore the	Y	This can only be done by Carestream authorized
DTBK-2	original device settings as provided by the manufacturer? Does the device have an integral data backup capability to	Yes	Service personnel
DTBK-3	removable media? Does the device have an integral data backup capability to remote	No	—
DTBK-4	storage?	No	
DTBK-5	Does the device have a backup capability for system configuration information, patch restoration, and software restoration?	Yes	
DTBK-6	Does the device provide the capability to check the integrity and authenticity of a backup?	No	
DIBK-0		NU	_
	EMERGENCY ACCESS (EMRG)		
	The ability of the device user to access personally identifiable information in case of a medical emergency situation that requires immediate access to stored personally identifiable information.		
	Does the device incorporate an emergency access (i.e. "break-	N-	
EMRG-1	glass") feature?	No	_
	HEALTH DATA INTEGRITY AND AUTHENTICITY (IGAU) How the device ensures that the stored data on the device has not been altered or destroyed in a non-authorized manner and is from the originator.		
	Does the device provide data integrity checking mechanisms of	No	
IGAU-1	stored health data (e.g., hash or digital signature)? Does the device provide error/failure protection and recovery	No	-
IGAU-2	mechanisms for stored health data (e.g., RAID-5)?	No	
	MALWARE DETECTION/PROTECTION (MLDP) The ability of the device to effectively prevent, detect and remove malicious software (malware).		
MLDP-1	Is the device capable of hosting executable software?	Yes	_

Manufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers AN8943 A Uncontrolled unless otherwise indicated

Carestream TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMAX Health. Inc. TX65 Laser Imager AN

Health, Inc.	TX65 Laser Imager	AN8943	1-Feb-2024
	Does the device support the use of anti-malware software (or other		
MLDP-2	anti-malware mechanism)? Provide details or reference in notes.	No	
MLDP-2.1	Does the device include anti-malware software by default?	No	_
MLDP-2.2	Does the device have anti-malware software available as an option?	No	_
	Does the device documentation allow the owner/operator to install		
MLDP-2.3	or update anti-malware software?	No	_
	Can the device owner/operator independently (re-)configure anti-		
MLDP-2.4	malware settings?	N/A	_
	Does notification of malware detection occur in the device user		
MLDP-2.5	interface?	N/A	
	Can only manufacturer-authorized persons repair systems when		
MLDP-2.6	malware has been detected?	Yes	
MLDP-2.7	Are malware notifications written to a log?	No	
	Are there any restrictions on anti-malware (e.g., purchase,		
MLDP-2.8	installation, configuration, scheduling)?	Yes	
	If the answer to MLDP-2 is NO, and anti-malware cannot be		
	installed on the device, are other compensating controls in place or		
MLDP-3	available?	Yes	_
	Does the device employ application whitelisting that restricts the		
MLDP-4	software and services that are permitted to be run on the device?	No	_
	Does the device employ a host-based intrusion		
MLDP-5	detection/prevention system?	No	-
	Can the host-based intrusion detection/prevention system be		
MLDP-5.1	configured by the customer?	N/A	_
	Can a host-based intrusion detection/prevention system be installed		
MLDP-5.2	by the customer?	No	
	NODE AUTHENTICATION (NAUT)		

No

No

The ability of the device to authenticate communication partners/nodes.

Does the device provide/support any means of node authentication that assures both the sender and the recipient of data are known to each other and are authorized to receive transferred information (e.g. Web APIs, SMTP, SNMP)?

NAUT-2	Are network access control mechanisms supported (E.g., does the device have an internal firewall, or use a network connection white list)?	Yes
NAUT-2.1	Is the firewall ruleset documented and available for review? Does the device use certificate-based network connection	No

CONNECTIVITY CAPABILITIES (CONN)

authentication?

NAUT-1

NAUT-3

	All network and removable media connections must be considered	
	in determining appropriate security controls. This section lists	
	connectivity capabilities that may be present on the device.	
CONN-1	Does the device have hardware connectivity capabilities?	Yes
CONN-1.1	Does the device support wireless connections?	No
CONN-1.1.1	Does the device support Wi-Fi?	No
CONN-1.1.2	Does the device support Bluetooth?	No
	Does the device support other wireless network connectivity (e.g.	
CONN-1.1.3	LTE, Zigbee, proprietary)?	No
	Does the device support other wireless connections (e.g., custom RF	
CONN-1.1.4	controls, wireless detectors)?	No
CONN-1.2	Does the device support physical connections?	Yes
CONN-1.2.1	Does the device have available RJ45 Ethernet ports?	Yes
CONN-1.2.2	Does the device have available USB ports?	No
	Does the device require, use, or support removable memory	
CONN-1.2.3	devices?	No
CONN-1.2.4	Does the device support other physical connectivity?	No
	Does the manufacturer provide a list of network ports and protocols	
CONN-2	that are used or may be used on the device?	Yes
	Can the device communicate with other systems within the	
CONN-3	customer environment?	Yes
	Can the device communicate with other systems external to the	
CONN-4	customer environment (e.g., a service host)?	Yes
CONN-5	Does the device make or receive API calls?	Yes
CONN-6	Does the device require an internet connection for its intended use?	No
CONN-7	Does the device support Transport Layer Security (TLS)?	No
CONN-7.1	Is TLS configurable?	No
	Does the device provide operator control functionality from a	
CONN-8	separate device (e.g., telemedicine)?	No
	Chatage of the Advised Device Council AdDCO for TDIAAAV TVAD TVEE	

_
_
_
_
USB ports are only accessible by Carestream
authorized Service personnel
This applies to remote connetivity If configured to
allow
 Public Classification

The system is protected from malicious attack by utilizing the OS firewall to block/disable all ports not needed for DICOM printing or secure remote access. Applicable OS Security Patches are incorporated into the software to provide additional protection

Manufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers AN8943 A Uncontrolled unless otherwise indicated PERSON AUTHENTICATION (PAUT)

PAUT-1

PAUT-1.1

PAUT-2 PAUT-3 PAUT-4 PAUT-5

PAUT-6 PAUT-7 PAUT-8 PAUT-9 PAUT-10 PAUT-11 PAUT-12 PAUT-13 PAUT-14 PAUT-14.1

thanticat ahility to nfic o tho dov

The ability to configure the device to authenticate users.		
Does the device support and enforce unique IDs and passwords for		
all users and roles (including service accounts)?	Yes	_
Does the device enforce authentication of unique IDs and		
passwords for all users and roles (including service accounts)?	Yes	_
Is the device configurable to authenticate users through an external		
authentication service (e.g., MS Active Directory, NDS, LDAP, OAuth,		
etc.)?	No	
Is the device configurable to lock out a user after a certain number		
of unsuccessful logon attempts?	No	_
Are all default accounts (e.g., technician service accounts,		
administrator accounts) listed in the documentation?	Yes	
Can all passwords be changed?	Yes	
Is the device configurable to enforce creation of user account		
passwords that meet established (organization specific) complexity		
rules?	No	
Does the device support account passwords that expire		
periodically?	No	
Does the device support multi-factor authentication?	Yes	For service level only
Does the device support single sign-on (SSO)?	No	
Can user accounts be disabled/locked on the device?	No	
Does the device support biometric controls?	No	
Does the device support physical tokens (e.g. badge access)?	No	_
Does the device support group authentication (e.g. hospital teams)?	No	_
Does the application or device store or manage authentication		
credentials?	No	

No

Yes

PHYSICAL LOCKS (PLOK)

Are credentials stored using a secure method?

Physical locks can prevent unauthorized users with physical access to the device from compromising the integrity and confidentiality of personally identifiable information stored on the device or on removable media Is the device software only? If yes, answer "N/A" to remaining PLOK-1 questions in this section. No Are all device components maintaining personally identifiable information (other than removable media) physically secure (i.e., PLOK-2 cannot remove without tools)? Yes Are all device components maintaining personally identifiable information (other than removable media) physically secured PLOK-3 behind an individually keyed locking device? No Does the device have an option for the customer to attach a PLOK-4 physical lock to restrict access to removable media? No

ROADMAP FOR THIRD PARTY COMPONENTS IN DEVICE LIFE

CYCLE (RDMP)

component end-of-life?

RDMP-4

Manufacturer's plans for security support of third-party

components within the device's life cycle.

RDMP-1	Was a secure software development process, such as ISO/IEC 27034 or IEC 62304, followed during product development?
	Does the manufacturer evaluate third-party applications and
	software components included in the device for secure
RDMP-2	development practices?
RDMP-3	Does the manufacturer maintain a web page or other source of information on software support dates and updates?
KDIVIP-5	Does the manufacturer have a plan for managing third-party
	boes the manufacturer have a pidit for managing third-party

Ļ			
	Yes	_	
	Yes	_	
	Yes		

SOFTWARE BILL OF MATERIALS (SBoM)

	A Software Bill of Material (SBoM) lists all the software			
	components that are incorporated into the device being described			
	for the purpose of operational security planning by the healthcare			
	delivery organization. This section supports controls in the RDMP			
	section.			
SBOM-1	Is the SBoM for this product available?	Yes		
	Does the SBoM follow a standard or common method in describing			
SBOM-2	software components?	Yes		
SBOM-2.1	Are the software components identified?	Yes		
	Are the developers/manufacturers of the software components			
SBOM-2.2	identified?	Yes		
	Are the major version numbers of the software components			
SBOM-2.3	identified?	Yes		
SBOM-2.4	Are any additional descriptive elements identified?	Yes		
Manufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers Publ			Classification	

ufacturer Disclosure Statement for Medical Device Security – MDS2 for TRIMAX TX40, TX55, and TX65 Laser Imagers Uncontrolled unless otherwise indicated AN8943 A

1-Feb-2024

Carestream Health, Inc.	TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMAX TX65 Laser Imager	(AN8943	1-Feb-2024
SBOM-3 SBOM-4	Does the device include a command or process method available to generate a list of software components installed on the device? Is there an update process for the SBoM?	No Yes	_
	SYSTEM AND APPLICATION HARDENING (SAHD)		
	The device's inherent resistance to cyber attacks and malware.		
SAHD-1	Is the device hardened in accordance with any industry standards?	Yes	_
SAHD-2	Has the device received any cybersecurity certifications?	No	-
SAHD-3	Does the device employ any mechanisms for software integrity checking Does the device employ any mechanism (e.g., release-specific hash	Yes	-
SAHD-3.1	key, checksums, digital signature, etc.) to ensure the installed software is manufacturer-authorized?	Yes	
	Does the device employ any mechanism (e.g., release-specific hash		_
SAHD-3.2	key, checksums, digital signature, etc.) to ensure the software updates are the manufacturer-authorized updates?	Yes	_
SAHD-4	Can the owner/operator perform software integrity checks (i.e., verify that the system has not been modified or tampered with)?	No	_
SAHD-5	Is the system configurable to allow the implementation of file-level, patient level, or other types of access controls?	No	_
SAHD-5.1	Does the device provide role-based access controls? Are any system or user accounts restricted or disabled by the	Yes	-
SAHD-6	manufacturer at system delivery? Are any system or user accounts configurable by the end user after	No	-
SAHD-6.1	initial configuration?	No	-
SAHD-6.2	Does this include restricting certain system or user accounts, such as service technicians, to least privileged access?	N/A	_
SAHD-7	Are all shared resources (e.g., file shares) which are not required for the intended use of the device disabled?	Yes	_
SAHD-8	Are all communication ports and protocols that are not required for the intended use of the device disabled?	Yes	_
	Are all services (e.g., telnet, file transfer protocol [FTP], internet		
SAHD-9	information server [IIS], etc.), which are not required for the intended use of the device deleted/disabled?	Yes	-
	Are all applications (COTS applications as well as OS-included applications, e.g., MS Internet Explorer, etc.) which are not required		
SAHD-10	for the intended use of the device deleted/disabled?	Yes	-
SAHD-11	Can the device prohibit boot from uncontrolled or removable media (i.e., a source other than an internal drive or memory component)? Can unauthorized software or hardware be installed on the device		-
SAHD-12	without the use of physical tools? Does the product documentation include information on	No	_
SAHD-13	operational network security scanning by users?	No	-
SAHD-14	Can the device be hardened beyond the default provided state?	No	_
SAHD-14.1	Are instructions available from vendor for increased hardening? Can the system prevent access to BIOS or other bootloaders during	N/A	
SHAD-15	boot? Have additional hardening methods not included in 2.3.19 been	No	
SAHD-16	used to harden the device?	No	_
	SECURITY GUIDANCE (SGUD)		
	Availability of security guidance for operator and administrator of the device and manufacturer sales and service.		
SGUD-1	Does the device include security documentation for the owner/operator?	No	-
SGUD-2	Does the device have the capability, and provide instructions, for the permanent deletion of data from the device or media?	No	_
SGUD-3	Are all access accounts documented?	No	-
SGUD-3.1	Can the owner/operator manage password control for all accounts? Does the product include documentation on recommended	No	-

HEALTH DATA STORAGE CONFIDENTIALITY (STCF)

compensating controls for the device?

SGUD-4

Can the owner/operator manage password control for all accounts? No Does the product include documentation on recommended

No

Carestream Health, Inc.	TRIMAX TX40 Laser Imager, TRIMAX TX55 Laser Imager, and TRIMAX TX65 Laser Imager	(AN8943	1-Feb-2024
	The ability of the device to ensure unauthorized access does not compromise the integrity and confidentiality of personally identifiable information stored on the device or removable media.		
STCF-1	Can the device encrypt data at rest?	No	No PII stored
STCF-1.1	Is all data encrypted or otherwise protected?	No	
STCF-1.2	Is the data encryption capability configured by default?	No	
CTCF 1 2		N1/A	
STCF-1.3 STCF-2	Are instructions available to the customer to configure encryption? Can the encryption keys be changed or configured?	N/A N/A	
STCF-3	Is the data stored in a database located on the device?	No	—
STCF-4	Is the data stored in a database external to the device?	No	
	TRANSMISSION CONFIDENTIALITY (TXCF)		
	The ability of the device to ensure the confidentiality of transmitted		
	personally identifiable information.		
	Can personally identifiable information be transmitted only via a		
TXCF-1	point-to-point dedicated cable?	No	_
	Is personally identifiable information encrypted prior to		
TXCF-2	transmission via a network or removable media?	No	—
TXCF-2.1	If data is not encrypted by default, can the customer configure encryption options?	Νο	
TACF=2.1	Is personally identifiable information transmission restricted to a		—
TXCF-3	fixed list of network destinations?	N/A	
TXCF-4	Are connections limited to authenticated systems?	No	—
	Are secure transmission methods supported/implemented (DICOM,		
TXCF-5	HL7, IEEE 11073)?	No	<u> </u>
	TRANSMISSION INTEGRITY (TXIG)		
	The ability of the device to ensure the integrity of transmitted data.		
	Does the device support any mechanism (e.g., digital signatures)		
TXIG-1	intended to ensure data is not modified during transmission?	No	
	Does the device include multiple sub-components connected by		—
TXIG-2	external cables?	No	
	REMOTE SERVICE (RMOT)		
	Remote service refers to all kinds of device maintenance activities		
	performed by a service person via network or other remote		
	connection.		
DMOT 1	Does the device permit remote service connections for device	¥	This is an exting it and he accepted
RMOT-1	analysis or repair? Does the device allow the owner/operator to initiative remote	Yes	This is an option, it can be restricted
RMOT-1.1	service sessions for device analysis or repair?	No	
			—
RMOT-1.2	Is there an indicator for an enabled and active remote session?	No	
	Can patient data be accessed or viewed from the device during the		Only if an image that contains annotations with ePHI
RMOT-1.3	remote session?	Yes	is viewed using the remote session
	Does the device permit or use remote service connections for		
RMOT-2	predictive maintenance data? Does the device have any other remotely accessible functionality	Yes	This is an option, it can be restricted
RMOT-3	(e.g. software updates, remote training)?	Yes	Software updates
	(e.g. sorenare apaates) remote training).		
	OTHER SECURITY CONSIDERATIONS (OTHR)		
	NONE		

Notes:

Note 1

Example note. Please keep individual notes to one cell. Please use separate notes for separate information